DEPARTMENT OF HEALTH & HUMAN SERVICES



New York District

Food & Drug Administration 300 Pearl Street, Suite 100 Buffalo, NY 14202

March 14, 2000

WARNING LETTER NYK 2000-49

CERTIFIED MAIL
RETURN RECEIPT REQUESTED

Yeongcheol Kim, M.D. 35-49 72nd Street Jackson Heights, New York 11372

RE: Facility ID: 150078

Dear Dr. Kim:

Your facility was inspected on March 6, 2000 by a representative of the New York City Bureau of Radiological Health, acting on behalf of the Food and Drug Administration. This inspection revealed serious regulatory problems involving the mammography at your facility.

Under a United States Federal law, the Mammography Quality Standards Act of 1992, your facility must meet specific requirements for mammography. These requirements help protect the health of women by assuring that a facility can perform quality mammography. The inspection revealed the following Level 1 findings at your facility:

- The medical physicist, did not have a Masters degree or higher in a physical science, with 20 semester hours in physics.
- The system to communicate results is not adequate because there is no system in place to provide timely lay summaries.

The specific problems noted above appeared on your MQSA Facility Inspection Report which was issued to your facility at the close of the inspection. These problems are identified as Level 1 because they identify failures to meet significant MQSA requirements.

Because these conditions may be symptomatic of serious underlying problems that could compromise the quality of mammography at your facility, they represent violations of the law which may result in FDA taking regulatory action without further notice to you. These actions include, but are not limited to, placing your facility under a Directed Plan of Correction; charging your facility for the cost of on-site monitoring; assessing civil money penalties up to \$10,000 for each failure to substantially comply with, or each day of failure to substantially comply with MQSA standards; suspension or revocation of your facility's FDA certificate; obtaining a court injunction against further mammography.

Yeongcheol Kim, M.D. Page 2

In addition, your response should address the Level 2 findings that were listed on the inspection report provided at the close of the inspection. The Level 2 findings are:

- There is no written procedure for handling consumer complaints.
- There is no written procedure for infection control.
- Corrective action for a failing image score (before further exams) was not documented for the mammography unit.
- Phantom QC is not adequate for the mammography unit because the image is not taken at the clinical setting.
- Seven of ten reports reviewed at random did not contain an assessment category.

It is necessary for you to act on these matters immediately. Please explain to this office in writing within fifteen (15) working days from the date that you receive this letter:

- 1. The specific steps you have taken to correct the violations noted in this letter. Please include supporting documentation such as new procedures.
- 2. Each step your facility is taking to prevent the recurrence of similar violations.

Please submit your response to the attention of Lisa M. Utz, Compliance Officer, U.S. Food and Drug Administration, 300 Pearl Street, Olympic Towers, Suite 100, Buffalo, New York 14202.

Finally, you should understand there are many FDA requirements pertaining to mammography. This letter pertains only to findings of your inspection and does not necessarily address other obligations you have under the law. You may obtain general information about all of FDA's requirements for mammography facilities by contacting the Mammography Quality Assurance Program, Food and Drug Administration, P.O. Box 6057, Columbia, Maryland 21045-6057 (1-800-838-7715), or through the Internet at http://www.fda.gov.

If you have any questions about mammography facility requirements, you may contact Murray L. Kurzman, Radiation Programs Manager, at (516) 921-2035.

Sincerely.

Brenda J. Holman District Director